



**DELAWARE HEALTH
AND SOCIAL SERVICES**

Division of Management Services

“DMS – Serving Those Who Serve Delaware”

PROCUREMENT

January 12, 2010

**ADDENDUM # 2
Questions & Answers**

BID# HSS 09-050

for

**Buy Equipment, Test Kits and Reagents for
Cystic Fibrosis Mutation Analysis**

Required for Use By

**Division of Public Health
Delaware Public Health Laboratory**

Date Due: **January 26, 2010, 11:00 AM** Local Time

ADDENDUM # 2

Please Note:

The attached sheets hereby become a part of the above mentioned bid.

Sylvia T. Adams
Procurement Coordinator
(302)255-9297

ADDENDUM # 2

Questions and Answers

BID # HSS 09-050

Buy Equipment, Test Kits and Reagents for
Cystic Fibrosis Mutation Analysis

1. What is the estimated monthly volumes to be run?

Initially volumes will be low as we will be running Delaware specimens only (20/month). Once other states join in, the volumes should increase dramatically. The rate of increase will depend on the ease/difficulty in getting those interstate contracts in place. My estimate at peak would be 100-150/month.

2. Does the New Born Screening Lab have a thermal cycler within the lab?

Yes, the Newborn Screening Lab has an Applied Biosystems GeneAmp PCR System 9700 thermocycler, brand new and never been used.